



# INSTRUCTIONS FOR DISPOSABLE INFUSION SETS

## INDICATIONS FOR USE

The product is gravity feed only and used to infusion.

### Main structure and properties

- Tubing and drip chamber are transparent, can easy to observe;
- Tubing is flexible enough to resist winding and twisting;
- The design of closure-piercing device to make it convenient for handle and usage;
- Flow regulator can make the flow rate change from minimum to maximum.

### USAGE

- Firstly, check valid period and completeness of package before using;
- Tear the package at the cut part and take out the infusion set;
- Close the regulator, pull out protective cap of closure-piercing device and push the spike into the disinfection closure of fluid container;
- Put fluid container upside down, press chamber and half of it was filled with injection solution; then open the regulator to expel air in tube.
- Penetrate vein according aseptic manipulation after expelling air then fix it.
- Observing and then change flow rate to normal.
- Infusion operation must have the professionals nursing to implement and guardian.
- 20 drops=(1±0.1)ml[(1±0.1)g]

### PRECAUTIONS

- It may causes hypothermia-esp, weak patients;
- It may causes abnormal thythem of ventricles of heart because of fast flow rate;
- There should maintain the guardianship during the infusion in case of any accident;
- It should be sterile and free from pyrogen before package opened;.
- It has a valid period of five years, do not use if it is beyond valid date.
- Discard after single use.

### WARNINGS

- Do not use if the package is opened or damaged, or protective cap is aparted, any part is missed, and foreign matter is found;
- It should be used by special trained person, and the skin should be correctly sterilized.
- The products can not be reused. The repetition use may cause cross infection, even serious illnesses such as hepatitis, HIV and other infectious disease.
- Always dispose of blood contaminated products in a manner consistent with established biohazard procedures.

### STORAGE

Product should be kept in a condition that relative humidity is less than 80%, it also should be free from corrosive gas, cool, dry with good ventilation and clean.

**Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.**

DO NOT REUSE	DO NOT USE IF PACKAGE IS DAMAGED	<b>STERILE</b> STERILE	<b>STERILE EO</b> STERILIZED USING ETHYLENE OXIDE
<b>MD</b> MEDICAL DEVICE	<b>UDI</b> UNIQUE DEVICE IDENTIFIER	CAUTION	<b>CE</b> 2862
Single sterile barrier system	Single sterile barrier system with protective packaging outside	REV.: B/0 EFFECTIVE: 2023-07-10	